

Urologic Guidewire Precaution and Warning

- 1) Please read the manual before use to understand the usage method of the device.
- 2) The device is sterilized by ethylene oxide, it's for single use only, and prohibited for re-sterilized.
- 3) Do not use the device if the package is damaged, contaminated or wet.
- 4) The device should only be used by physicians who have comprehensive knowledge of the indications, techniques and risks of the procedure.
- 5) Do not use the device after its expiry date.
- 6) The device should be disposed as medical waste after use.